

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3262

Name: RITTER 1  
Sterilizer ID: V2178198  
05/19/2023 10:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.9 F 0.0  
2:00 163.4 F 0.0  
4:00 185.1 F 2.4  
6:00 202.8 F 6.9  
8:00 217.9 F 11.7  
10:00 224.8 F 6.3  
12:00 237.6 F 11.5  
14:00 236.3 F 9.0  
16:00 239.4 F 9.6  
18:00 249.1 F 14.5  
20:00 261.6 F 21.5  
22:00 272.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 27.9  
1:00 272.5 F 28.5  
2:00 271.5 F 27.9  
3:00 272.7 F 28.6  
4:00 271.7 F 27.9

Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
5:00 217.7 F 0.0  
10:00 230.7 F 0.0  
15:00 227.0 F 0.0  
20:00 222.7 F 0.0  
25:00 219.7 F 0.0  
30:00 219.2 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 22:23  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:34  
CYCLE COMPLETE

05/19/2023 11:40 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_