
Midmark M11 - v1.0.5
Total Cycles: 3369

Name: RITTER2
Sterilizer ID: V2313228
02/08/2024 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.1 F 0.0
2:00 143.9 F 0.3
4:00 143.3 F 2.5
6:00 165.1 F 5.8
8:00 186.6 F 9.9
10:00 204.3 F 8.4
12:00 215.1 F 7.8
14:00 228.3 F 12.4
16:00 234.3 F 9.4
18:00 238.5 F 10.0
20:00 246.2 F 13.8
22:00 257.8 F 20.0
24:00 268.1 F 26.5
26:00 272.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.9 F 29.7
2:00 272.1 F 29.9
3:00 272.4 F 29.6
4:00 272.4 F 30.0

Min 271.7 F 29.3
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 226.4 F 0.0
10:00 230.8 F 0.0
15:00 223.0 F 0.0
20:00 218.5 F 0.0
25:00 216.8 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:08
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:04:36
CYCLE COMPLETE

02/08/2024 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____