
Midmark M11 - v1.0.5
Total Cycles: 4003

Name: RITTER2
Sterilizer ID: V2313228
10/04/2024 04:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.5 F 0.0
2:00 122.3 F 0.0
4:00 144.1 F 1.0
6:00 173.7 F 4.9
8:00 197.9 F 9.2
10:00 211.9 F 9.9
12:00 224.6 F 8.3
14:00 236.0 F 13.2
16:00 236.7 F 10.3
18:00 236.0 F 8.5
20:00 248.5 F 14.9
22:00 260.0 F 21.3
24:00 270.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.6
2:00 271.4 F 28.8
3:00 272.6 F 29.7
4:00 271.9 F 28.9

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 211.4 F 0.0
10:00 197.3 F 0.0
15:00 193.2 F 0.0
20:00 190.6 F 0.0
25:00 190.5 F 0.0
30:00 207.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:49
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:03:12
CYCLE COMPLETE

10/04/2024 05:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____