
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006011
11:31 19/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:44
PRESSURIZING 1:50
STERILIZING 2:52
136.5 C 222kPa 2:52
136.4 C 225kPa 3:22
136.4 C 223kPa 3:52
136.2 C 222kPa 4:22
136.1 C 225kPa 4:52
136.4 C 218kPa 5:22
135.9 C 220kPa 5:52
136.3 C 219kPa 6:22
136.3 C 219kPa 6:22
VENTING 6:22
AIR DRYING 7:14
CYCLE COMPLETE 10:35

Digital Signature #
0B4242837B36510F
