
Midmark M11 - v1.0.5
Total Cycles: 3484

Name: RITTER2
Sterilizer ID: V2313228
03/27/2024 04:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.9 F 0.0
2:00 126.0 F 0.6
4:00 146.4 F 3.0
6:00 170.4 F 6.7
8:00 193.9 F 10.6
10:00 213.0 F 6.2
12:00 229.8 F 8.6
14:00 237.8 F 13.2
16:00 237.2 F 10.3
18:00 236.0 F 8.9
20:00 248.0 F 14.8
22:00 259.3 F 21.0
24:00 269.8 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.6
1:00 271.7 F 29.6
2:00 271.6 F 29.5
3:00 272.1 F 29.6
4:00 272.8 F 29.8

Min 271.4 F 29.5
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.6 F 0.0
5:00 234.4 F 0.0
10:00 227.1 F 0.0
15:00 229.8 F 0.0
20:00 222.6 F 0.0
25:00 205.8 F 0.0
30:00 182.9 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 25:38
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:04:28
CYCLE COMPLETE

03/27/2024 05:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____