

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2119

Name: RITTER 1  
Sterilizer ID: V2178198  
07/13/2022 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
2:00 185.1 F 0.8  
4:00 194.5 F 4.8  
6:00 209.2 F 10.0  
8:00 224.6 F 5.9  
10:00 232.0 F 10.3  
12:00 234.6 F 7.7  
14:00 243.5 F 13.1  
16:00 242.7 F 11.5  
18:00 255.3 F 18.1  
20:00 266.1 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.2 F 29.3  
2:00 272.0 F 28.8  
3:00 272.9 F 29.5  
4:00 271.7 F 28.7

Min 270.9 F 28.1  
Max 273.1 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 232.0 F 0.0  
10:00 224.3 F 0.0  
15:00 229.3 F 0.0  
20:00 225.3 F 0.0  
25:00 214.9 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:59  
CYCLE COMPLETE

07/13/2022 10:33 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_