
Midmark M11 - v1.0.5
Total Cycles: 2572

Name: RITTER2
Sterilizer ID: V2313228
06/27/2023 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 180.8 F 2.4
4:00 190.2 F 6.7
6:00 214.1 F 11.9
8:00 222.5 F 7.4
10:00 236.2 F 12.8
12:00 237.3 F 10.6
14:00 235.8 F 9.1
16:00 250.2 F 15.9
18:00 261.9 F 22.7
20:00 272.6 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.6 F 29.9
2:00 272.0 F 29.3
3:00 271.4 F 29.3
4:00 272.6 F 29.9

Min 271.3 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 207.6 F 0.0
10:00 216.0 F 0.0
15:00 221.1 F 0.0
20:00 221.2 F 0.0
25:00 219.9 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

06/27/2023 09:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____