

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2642

Name: RITTER 1  
Sterilizer ID: V2178198  
12/05/2022 07:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.5 F 0.0  
2:00 160.0 F 2.4  
4:00 183.8 F 6.0  
6:00 202.0 F 10.8  
8:00 216.3 F 4.8  
10:00 228.4 F 10.2  
12:00 231.9 F 7.6  
14:00 243.2 F 12.9  
16:00 242.0 F 11.6  
18:00 255.4 F 18.5  
20:00 267.5 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.3 F 28.9  
2:00 271.9 F 29.0  
3:00 272.0 F 28.8  
4:00 271.9 F 29.0  
  
Min 271.4 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.4 F 0.0  
5:00 222.7 F 0.0  
10:00 235.9 F 0.0  
15:00 232.6 F 0.0  
20:00 224.4 F 0.0  
25:00 222.9 F 0.0  
30:00 221.9 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 21:21  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:40  
CYCLE COMPLETE

12/05/2022 08:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_