
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006563
14:31 06/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:27
PRESSURIZING 3:28
STERILIZING 4:29
136.4 C 224kPa 4:29
136.1 C 221kPa 4:59
136.4 C 225kPa 5:29
136.4 C 224kPa 5:59
136.4 C 225kPa 6:29
136.4 C 223kPa 6:59
136.3 C 223kPa 7:29
136.4 C 223kPa 7:59
136.4 C 223kPa 7:59
VENTING 7:59
AIR DRYING 8:54
CYCLE COMPLETE 8:56

Digital Signature #
2092C073CB113F7C
