

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2602

Name: RITTER 1  
Sterilizer ID: V2178198  
11/23/2022 10:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
2:00 204.6 F 4.0  
4:00 196.8 F 9.2  
6:00 220.9 F 7.9  
8:00 233.9 F 10.4  
10:00 234.0 F 7.2  
12:00 244.3 F 12.3  
14:00 243.6 F 12.4  
16:00 256.5 F 19.3  
18:00 267.8 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.7 F 29.5  
2:00 272.1 F 29.0  
3:00 271.2 F 28.8  
4:00 272.5 F 29.6  
  
Min 271.0 F 28.6  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 241.4 F 0.0  
10:00 229.9 F 0.0  
15:00 231.2 F 0.0  
20:00 232.4 F 0.0  
25:00 228.3 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 19:17  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:08  
CYCLE COMPLETE

11/23/2022 11:54 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_