
Midmark M11 - v1.0.0
Total Cycles: 5200

Name: RITTER 1
Sterilizer ID: V2178198
12/18/2024 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 73.5 F 0.0
4:00 88.0 F 0.0
6:00 122.4 F 0.0
8:00 151.5 F 0.1
10:00 178.5 F 0.1
12:00 198.1 F 0.4
14:00 210.8 F 1.6
16:00 223.6 F 5.0
18:00 231.0 F 8.0
20:00 238.8 F 11.6
22:00 234.0 F 8.0
24:00 246.6 F 12.9
26:00 244.1 F 12.0
28:00 241.3 F 10.7
30:00 254.6 F 17.3
32:00 265.4 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 27.9
1:00 272.4 F 28.5
2:00 272.8 F 28.5
3:00 272.1 F 28.0
4:00 271.7 F 27.9

Min 271.6 F 27.8
Max 272.9 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 170.0 F 0.0
10:00 188.1 F 0.0
15:00 211.4 F 0.0
20:00 229.3 F 0.0
25:00 229.3 F 0.0
30:00 225.2 F 0.0

Drying Complete

FILLING: 0:40
HEATING: 33:57
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:12:07
CYCLE COMPLETE

12/18/2024 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____