
Midmark M11 - v1.0.0
Total Cycles: 919

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2021 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.7 F 0.0
2:00 114.6 F 0.0
4:00 129.1 F 0.6
6:00 157.5 F 2.9
8:00 181.8 F 6.0
10:00 203.1 F 10.4
12:00 215.7 F 5.3
14:00 224.8 F 8.1
16:00 236.1 F 12.7
18:00 236.0 F 9.6
20:00 238.4 F 9.7
22:00 246.3 F 13.5
24:00 257.7 F 19.7
26:00 268.1 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.2
2:00 271.8 F 28.5
3:00 272.0 F 29.0
4:00 272.4 F 29.1

Min 271.4 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 224.8 F 0.0
10:00 222.6 F 0.0
15:00 228.3 F 0.0
20:00 227.9 F 0.0
25:00 230.8 F 0.0
30:00 185.9 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 27:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:05:32
CYCLE COMPLETE

06/21/2021 09:45 AM
ACCEPTED: X
REJECTED:
Operator: U01