
Midmark M11 - v1.0.5
Total Cycles: 830

Name: RITTER2
Sterilizer ID: V2313228
02/23/2022 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.2 F 0.0
2:00 182.1 F 4.0
4:00 206.4 F 9.5
6:00 221.1 F 7.5
8:00 233.2 F 11.3
10:00 235.1 F 9.2
12:00 238.1 F 9.5
14:00 249.7 F 15.4
16:00 262.0 F 22.5
18:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.6 F 29.5
2:00 271.4 F 28.8
3:00 272.7 F 29.6
4:00 271.9 F 29.0

Min 271.3 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 233.3 F 0.0
10:00 229.4 F 0.0
15:00 224.6 F 0.0
20:00 220.1 F 0.0
25:00 226.5 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 2:52
HEATING: 18:19
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

02/23/2022 04:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____