
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001406
13:18 27/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:00
STERILIZING 4:03
136.0 C 220kPa 4:03
136.0 C 222kPa 4:33
136.4 C 224kPa 5:03
136.0 C 221kPa 5:33
136.3 C 224kPa 6:03
136.3 C 223kPa 6:33
136.1 C 222kPa 7:03
136.3 C 224kPa 7:33
136.3 C 224kPa 7:33
VENTING 7:33
AIR DRYING 8:41
CYCLE COMPLETE 18:30

Digital Signature #
4B0CA49D20FFDECF
