
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002955
8:43 13/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:29
STERILIZING 4:40
136.0 C 220kPa 4:40
136.4 C 224kPa 5:10
136.0 C 220kPa 5:40
136.0 C 222kPa 6:10
136.4 C 224kPa 6:40
136.4 C 223kPa 7:10
136.4 C 222kPa 7:40
136.4 C 223kPa 8:10
136.4 C 223kPa 8:10
VENTING 8:10
AIR DRYING 9:03
CYCLE COMPLETE 13:32

Digital Signature #
A255082F98650977
