
Midmark M11 - v1.0.0
Total Cycles: 4587

Name: RITTER 1
Sterilizer ID: V2178198
05/28/2024 07:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 71.8 F 0.0
4:00 84.3 F 0.1
6:00 117.7 F 0.4
8:00 146.9 F 1.8
10:00 171.9 F 5.0
12:00 189.1 F 8.4
14:00 204.1 F 11.8
16:00 215.3 F 3.9
18:00 223.3 F 7.8
20:00 233.3 F 11.7
22:00 231.8 F 7.6
24:00 244.0 F 13.0
26:00 241.5 F 11.2
28:00 253.5 F 17.2
30:00 264.1 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.0 F 28.8
2:00 272.4 F 28.8
3:00 271.2 F 28.3
4:00 272.6 F 29.2

Min 271.2 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 222.3 F 0.0
10:00 224.3 F 0.0
15:00 235.6 F 0.0
20:00 222.2 F 0.0
25:00 216.9 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:59
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:10:15
CYCLE COMPLETE

05/28/2024 09:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____