

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5154

Name: RITTER 1  
Sterilizer ID: V2178198  
12/04/2024 09:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.9 F 0.0  
2:00 71.4 F 0.0  
4:00 80.0 F 0.0  
6:00 113.5 F 0.0  
8:00 146.3 F 0.3  
10:00 175.6 F 0.6  
12:00 197.6 F 2.1  
14:00 220.0 F 5.9  
16:00 226.4 F 9.4  
18:00 236.4 F 13.4  
20:00 236.0 F 9.0  
22:00 241.5 F 11.0  
24:00 243.9 F 12.2  
26:00 241.2 F 10.8  
28:00 254.5 F 17.5  
30:00 265.8 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.0  
1:00 271.8 F 28.3  
2:00 272.8 F 28.8  
3:00 272.3 F 28.3  
4:00 271.6 F 27.9

Min 271.5 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
5:00 205.9 F 0.0  
10:00 182.2 F 0.0  
15:00 186.3 F 0.0  
20:00 188.8 F 0.0  
25:00 208.6 F 0.0  
30:00 212.9 F 0.0  
Drying Complete

FILLING: 0:41  
HEATING: 31:40  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:09:48  
CYCLE COMPLETE

12/04/2024 10:49 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_