

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1859

Name: RITTER2  
Sterilizer ID: V2313228  
12/19/2022 04:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.3 F 0.0  
2:00 191.2 F 4.5  
4:00 205.4 F 10.2  
6:00 218.6 F 6.0  
8:00 234.2 F 12.2  
10:00 238.0 F 10.5  
12:00 237.3 F 9.6  
14:00 252.5 F 17.0  
16:00 264.5 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.2 F 29.8  
2:00 272.9 F 29.8  
3:00 272.4 F 29.4  
4:00 271.9 F 29.1  
  
Min 271.2 F 28.8  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.7 F 0.0  
5:00 236.6 F 0.0  
10:00 227.9 F 0.0  
15:00 230.7 F 0.0  
20:00 226.5 F 0.0  
25:00 229.1 F 0.0  
30:00 229.7 F 0.0  
Drying Complete

FILLING: 3:12  
HEATING: 17:55  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:17  
CYCLE COMPLETE

12/19/2022 05:45 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_