
Midmark M11 - v1.0.5
Total Cycles: 326

Name: RITTER2
Sterilizer ID: V2313228
08/31/2021 12:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.2 F 0.0
2:00 179.7 F 2.8
4:00 179.9 F 7.0
6:00 199.5 F 12.1
8:00 218.3 F 7.3
10:00 228.5 F 12.7
12:00 240.1 F 10.8
14:00 237.9 F 9.7
16:00 251.2 F 16.3
18:00 262.8 F 22.9
20:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 272.5 F 30.4
2:00 274.5 F 31.0
3:00 273.4 F 30.8
4:00 273.0 F 30.6

Min 270.9 F 29.6
Max 274.5 F 31.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 231.9 F 0.0
10:00 228.9 F 0.0
15:00 232.0 F 0.0
20:00 226.8 F 0.0
25:00 228.9 F 0.0
30:00 231.5 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

08/31/2021 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____