
Midmark M11 - v1.0.0
Total Cycles: 3255

Name: RITTER 1
Sterilizer ID: V2178198
05/18/2023 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 213.5 F 0.0
2:00 207.6 F 0.0
4:00 209.8 F 4.7
6:00 219.1 F 10.0
8:00 230.1 F 7.0
10:00 237.2 F 11.3
12:00 237.6 F 9.0
14:00 239.5 F 9.3
16:00 249.3 F 14.7
18:00 261.3 F 21.6
20:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.7 F 29.1
2:00 272.1 F 28.4
3:00 271.9 F 28.5
4:00 272.8 F 29.1

Min 271.6 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 218.2 F 0.0
10:00 232.2 F 0.0
15:00 243.9 F 0.0
20:00 241.0 F 0.0
25:00 229.9 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

05/18/2023 11:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____