
Midmark M11 - v1.0.0
Total Cycles: 2160

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2022 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.9 F 0.0
2:00 125.0 F 0.0
4:00 141.9 F 2.2
6:00 171.2 F 5.8
8:00 190.0 F 9.8
10:00 213.6 F 7.1
12:00 225.7 F 7.6
14:00 237.9 F 12.1
16:00 233.3 F 8.5
18:00 241.9 F 11.5
20:00 243.7 F 12.2
22:00 255.6 F 18.5
24:00 266.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.5 F 29.2
2:00 271.7 F 28.5
3:00 271.6 F 28.9
4:00 272.5 F 29.1

Min 271.2 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.9 F 0.0
5:00 209.6 F 0.0
10:00 174.1 F 0.0
15:00 155.0 F 0.0
20:00 150.5 F 0.0
25:00 146.8 F 0.0
30:00 143.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:40
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:03:46
CYCLE COMPLETE

07/21/2022 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____