
Midmark M11 - v1.0.0
Total Cycles: 762

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2021 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.5 F 0.0
2:00 173.9 F 2.7
4:00 175.6 F 7.2
6:00 207.9 F 12.4
8:00 220.7 F 8.1
10:00 239.6 F 10.9
12:00 239.5 F 11.2
14:00 238.0 F 9.3
16:00 252.3 F 16.3
18:00 264.1 F 23.5
20:00 270.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.4 F 29.4
2:00 272.2 F 29.2
3:00 271.8 F 29.0
4:00 271.3 F 28.8

Min 271.2 F 28.8
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 239.3 F 0.0
10:00 186.8 F 0.0
15:00 173.7 F 0.0
20:00 164.8 F 0.0
Dry Cycle Aborted

FILLING: 1:12
HEATING: 20:29
STERILIZING: 4:00
VENTING: 2:57
DRYING: 21:23
TOTAL CYCLE: 00:50:01

03/31/2021 02:22 PM
ACCEPTED: X
REJECTED:
Operator: U01