
Midmark M11 - v1.0.0
Total Cycles: 1806

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2022 07:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.8 F 0.0
2:00 73.7 F 0.0
4:00 86.2 F 0.0
6:00 119.1 F 0.1
8:00 154.5 F 0.8
10:00 170.8 F 3.9
12:00 191.3 F 7.1
14:00 211.5 F 10.4
16:00 222.6 F 8.1
18:00 224.0 F 6.5
20:00 233.3 F 10.5
22:00 235.9 F 8.4
24:00 243.3 F 12.0
26:00 239.9 F 10.0
28:00 252.4 F 16.2
30:00 263.1 F 22.5
32:00 272.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 28.6
2:00 272.4 F 29.0
3:00 272.1 F 28.4
4:00 272.8 F 29.0

Min 271.5 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 223.1 F 0.0
10:00 225.1 F 0.0
15:00 200.1 F 0.0
20:00 214.4 F 0.0
25:00 218.5 F 0.0
30:00 228.2 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:58
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:11:03
CYCLE COMPLETE

04/11/2022 08:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____