
Midmark M11 - v1.0.5
Total Cycles: 2593

Name: RITTER2
Sterilizer ID: V2313228
06/30/2023 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.4 F 0.0
2:00 207.1 F 2.2
4:00 206.3 F 7.2
6:00 218.7 F 13.0
8:00 230.2 F 8.8
10:00 239.5 F 10.3
12:00 242.4 F 12.6
14:00 241.9 F 11.6
16:00 255.2 F 18.5
18:00 266.7 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.6
2:00 271.8 F 28.8
3:00 271.8 F 29.4
4:00 272.3 F 29.4

Min 271.2 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 223.1 F 0.0
10:00 234.7 F 0.0
15:00 226.1 F 0.0
20:00 224.5 F 0.0
25:00 222.5 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

06/30/2023 09:55 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____