
Midmark M11 - v1.0.0
Total Cycles: 3166

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2023 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.3 F 0.0
2:00 164.5 F 0.0
4:00 168.2 F 1.6
6:00 195.4 F 6.1
8:00 213.5 F 11.1
10:00 223.5 F 5.9
12:00 234.7 F 11.2
14:00 233.8 F 8.1
16:00 242.7 F 11.3
18:00 245.9 F 13.0
20:00 258.3 F 19.7
22:00 269.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.8
2:00 271.6 F 28.2
3:00 272.7 F 29.0
4:00 271.8 F 28.2

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 214.9 F 0.0
10:00 234.4 F 0.0
15:00 228.2 F 0.0
20:00 225.3 F 0.0
25:00 227.5 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:02
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

04/27/2023 09:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____