

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1920

Name: RITTER 1  
Sterilizer ID: V2178198  
05/09/2022 11:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
2:00 178.1 F 3.5  
4:00 193.4 F 8.1  
6:00 216.1 F 13.4  
8:00 225.9 F 8.7  
10:00 236.4 F 10.6  
12:00 240.9 F 11.3  
14:00 238.5 F 9.6  
16:00 252.5 F 16.5  
18:00 264.3 F 23.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.6 F 28.9  
2:00 271.6 F 28.1  
3:00 272.3 F 28.9  
4:00 272.3 F 28.6  
  
Min 271.3 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
5:00 212.5 F 0.0  
10:00 206.6 F 0.0  
15:00 205.9 F 0.0  
20:00 213.0 F 0.0  
25:00 243.1 F 0.0  
30:00 228.1 F 0.0  
  
Drying Complete

FILLING: 1:34  
HEATING: 19:58  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:41  
CYCLE COMPLETE

05/09/2022 12:50 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02