

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2457

Name: RITTER2  
Sterilizer ID: V2313228  
06/02/2023 10:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 124.8 F 0.0  
2:00 125.3 F 0.0  
4:00 130.4 F 0.2  
6:00 161.4 F 3.0  
8:00 177.1 F 6.5  
10:00 196.3 F 10.3  
12:00 216.1 F 7.8  
14:00 224.3 F 8.7  
16:00 233.0 F 13.3  
18:00 236.7 F 10.2  
20:00 236.6 F 9.1  
22:00 247.2 F 14.4  
24:00 258.4 F 20.4  
26:00 268.4 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.0  
1:00 272.5 F 29.9  
2:00 272.7 F 29.7  
3:00 272.4 F 30.0  
4:00 272.0 F 29.4

Min 271.0 F 28.9  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 230.7 F 0.0  
10:00 212.1 F 0.0  
15:00 191.8 F 0.0  
20:00 185.6 F 0.0  
25:00 189.2 F 0.0  
30:00 196.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 27:20  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:05:41  
CYCLE COMPLETE

06/02/2023 11:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_