
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008042
9:14 08/11/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:24
STERILIZING 3:28
136.5 C 222kPa 3:28
136.4 C 223kPa 3:58
136.4 C 217kPa 4:28
136.2 C 221kPa 4:58
136.1 C 221kPa 5:28
136.2 C 222kPa 5:58
136.3 C 221kPa 6:28
136.4 C 223kPa 6:58
136.4 C 223kPa 6:58
VENTING 6:58
AIR DRYING 7:50
CYCLE COMPLETE 25:09

Digital Signature #
7CDDBBEB5115DB44
