
Midmark M11 - v1.0.0
Total Cycles: 1703

Name: RITTER 1
Sterilizer ID: V2178198
03/09/2022 02:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.1 F 0.0
2:00 187.3 F 2.9
4:00 197.0 F 7.2
6:00 213.1 F 12.4
8:00 225.5 F 7.1
10:00 237.1 F 12.3
12:00 236.5 F 9.5
14:00 237.2 F 8.8
16:00 248.7 F 14.6
18:00 260.5 F 21.3
20:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.0
2:00 271.2 F 28.4
3:00 272.6 F 29.1
4:00 271.5 F 28.4

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 237.6 F 0.0
10:00 226.9 F 0.0
15:00 232.4 F 0.0
20:00 225.3 F 0.0
25:00 229.7 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 20:37
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

03/09/2022 03:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____