
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000942
11:20 18/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:58
STERILIZING 4:12
136.0 C 225kPa 4:12
136.2 C 231kPa 4:42
136.2 C 231kPa 5:12
136.1 C 230kPa 5:42
136.4 C 232kPa 6:12
136.4 C 233kPa 6:42
136.4 C 233kPa 7:12
136.2 C 230kPa 7:42
136.2 C 230kPa 7:42
VENTING 7:42
AIR DRYING 8:33
CYCLE COMPLETE 23:33

Digital Signature #
1AC457318E0F800F
