

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3534

Name: RITTER 1  
Sterilizer ID: V2178198  
07/17/2023 11:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
2:00 174.6 F 1.5  
4:00 190.2 F 4.6  
6:00 209.9 F 8.9  
8:00 220.3 F 13.4  
10:00 227.1 F 7.6  
12:00 238.9 F 12.1  
14:00 235.9 F 9.2  
16:00 240.7 F 10.3  
18:00 246.2 F 13.3  
20:00 257.5 F 19.4  
22:00 268.2 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.7 F 29.1  
2:00 271.6 F 28.4  
3:00 271.7 F 28.8  
4:00 272.4 F 29.0

Min 271.1 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.8 F 0.0  
5:00 234.5 F 0.0  
10:00 230.2 F 0.0  
15:00 226.2 F 0.0  
20:00 223.4 F 0.0  
25:00 219.7 F 0.0  
30:00 216.2 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 23:17  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:02:08  
CYCLE COMPLETE

07/17/2023 12:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_