
Midmark M11 - v1.0.5
Total Cycles: 4408

Name: RITTER2
Sterilizer ID: V2313228
03/06/2025 04:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.6 F 0.0
2:00 118.6 F 1.0
4:00 125.9 F 3.3
6:00 156.1 F 6.7
8:00 182.3 F 10.3
10:00 206.6 F 5.9
12:00 219.9 F 7.7
14:00 228.4 F 11.9
16:00 229.8 F 8.6
18:00 241.2 F 12.7
20:00 240.7 F 12.1
22:00 251.9 F 17.9
24:00 263.0 F 24.1
26:00 272.0 F 30.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 30.1
1:00 272.2 F 31.3
2:00 272.3 F 31.3
3:00 272.1 F 31.1
4:00 271.9 F 30.8

Min 270.6 F 29.9
Max 273.3 F 31.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 227.8 F 0.0
10:00 224.5 F 0.0
15:00 217.2 F 0.0
20:00 209.1 F 0.0
25:00 201.8 F 0.0
30:00 197.7 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:31
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:05:22
CYCLE COMPLETE

03/06/2025 05:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____