

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2931

Name: RITTER 1  
Sterilizer ID: V2178198  
02/22/2023 02:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.9 F 0.0  
2:00 186.1 F 4.5  
4:00 203.5 F 9.5  
6:00 221.1 F 6.6  
8:00 231.1 F 10.7  
10:00 233.9 F 7.8  
12:00 241.8 F 11.0  
14:00 246.8 F 13.5  
16:00 259.7 F 20.6  
18:00 271.4 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.5 F 28.9  
2:00 271.4 F 28.1  
3:00 272.2 F 28.9  
4:00 272.2 F 28.7  
  
Min 271.1 F 28.0  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.8 F 0.0  
5:00 224.0 F 0.0  
10:00 228.1 F 0.0  
15:00 228.5 F 0.0  
20:00 228.9 F 0.0  
25:00 228.7 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 18:40  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:57:53  
CYCLE COMPLETE

02/22/2023 03:13 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_