
Midmark M11 - v1.0.0
Total Cycles: 4455

Name: RITTER 1
Sterilizer ID: V2178198
04/15/2024 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.0 F 0.0
2:00 68.3 F 0.2
4:00 75.5 F 1.0
6:00 100.7 F 2.8
8:00 127.1 F 5.3
10:00 154.9 F 8.3
12:00 177.5 F 11.6
14:00 200.9 F 1.9
16:00 207.7 F 5.0
18:00 218.0 F 8.1
20:00 228.0 F 11.6
22:00 227.8 F 6.6
24:00 239.3 F 11.5
26:00 235.6 F 8.7
28:00 248.2 F 14.4
30:00 258.6 F 20.1
32:00 268.3 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.1
2:00 271.7 F 28.4
3:00 271.8 F 28.8
4:00 272.4 F 28.9

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 206.2 F 0.0
10:00 205.4 F 0.0
15:00 190.7 F 0.0
20:00 226.1 F 0.0
25:00 230.9 F 0.0
30:00 230.0 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:20
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:11:40
CYCLE COMPLETE

04/15/2024 09:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____