
Midmark M11 - v1.0.5
Total Cycles: 3788

Name: RITTER2
Sterilizer ID: V2313228
07/16/2024 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.9 F 0.0
2:00 171.6 F 2.4
4:00 178.5 F 6.0
6:00 197.7 F 10.1
8:00 213.4 F 7.5
10:00 223.3 F 7.6
12:00 234.1 F 12.1
14:00 234.8 F 9.4
16:00 239.2 F 10.4
18:00 246.3 F 14.0
20:00 257.3 F 19.9
22:00 267.7 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.7 F 29.9
2:00 271.8 F 29.1
3:00 271.8 F 29.7
4:00 272.4 F 29.7

Min 271.2 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 210.4 F 0.0
10:00 207.9 F 0.0
15:00 202.3 F 0.0
20:00 193.2 F 0.0
25:00 200.6 F 0.0
30:00 198.5 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 23:27
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:02:41
CYCLE COMPLETE

07/16/2024 12:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____