

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4613

Name: RITTER 1  
Sterilizer ID: V2178198  
06/06/2024 10:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.0 F 0.0  
2:00 123.9 F 0.7  
4:00 127.5 F 2.7  
6:00 145.1 F 5.4  
8:00 168.5 F 8.6  
10:00 191.2 F 12.2  
12:00 208.6 F 4.2  
14:00 218.2 F 8.0  
16:00 229.6 F 12.1  
18:00 230.3 F 7.9  
20:00 241.4 F 12.9  
22:00 239.9 F 10.8  
24:00 251.3 F 16.4  
26:00 261.6 F 22.3  
28:00 271.3 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.9 F 29.8  
2:00 271.9 F 29.1  
3:00 272.4 F 29.8  
4:00 272.5 F 29.5

Min 271.6 F 29.1  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 209.0 F 0.0  
10:00 203.7 F 0.0  
15:00 212.6 F 0.0  
20:00 225.5 F 0.0  
25:00 236.0 F 0.0  
30:00 224.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:33  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:08:14  
CYCLE COMPLETE

06/06/2024 11:29 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_