

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002233  
11:39 07/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:32  
PRESSURIZING 1:41  
STERILIZING 2:32  
136.0 C 222kPa 2:32  
136.1 C 223kPa 3:02  
136.4 C 226kPa 3:32  
136.5 C 223kPa 4:02  
136.4 C 226kPa 4:32  
136.4 C 222kPa 5:02  
136.3 C 224kPa 5:32  
136.4 C 224kPa 6:02  
136.4 C 224kPa 6:02  
VENTING 6:02  
AIR DRYING 6:54  
CYCLE COMPLETE 36:55  
-----

Digital Signature #  
F4960B547D670C7F  
-----