

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4786

Name: RITTER 1  
Sterilizer ID: V2178198  
08/07/2024 07:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.6 F 0.0  
2:00 69.1 F 0.0  
4:00 86.6 F 0.0  
6:00 127.8 F 0.1  
8:00 156.8 F 0.9  
10:00 183.8 F 2.3  
12:00 204.7 F 5.6  
14:00 214.6 F 9.0  
16:00 226.8 F 12.6  
18:00 225.3 F 6.5  
20:00 237.4 F 10.9  
22:00 232.7 F 7.1  
24:00 246.1 F 13.4  
26:00 243.9 F 11.9  
28:00 256.4 F 18.4  
30:00 267.6 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.2 F 28.2  
2:00 272.7 F 28.7  
3:00 271.6 F 28.0  
4:00 272.3 F 28.3

Min 271.5 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.0 F 0.0  
5:00 210.0 F 0.0  
10:00 223.6 F 0.0  
15:00 222.0 F 0.0  
20:00 221.1 F 0.0  
25:00 232.5 F 0.0  
30:00 231.0 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 31:18  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:09:20  
CYCLE COMPLETE

08/07/2024 08:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_