
Midmark M11 - v1.0.0
Total Cycles: 633

Name: MIDMARK
Sterilizer ID: V2178198
01/27/2021 07:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

| mm:ss | Degrees | PSI |
|-------|---------|------|
| 0:00 | 71.0 F | 0.0 |
| 2:00 | 74.3 F | 0.0 |
| 4:00 | 87.4 F | 0.1 |
| 6:00 | 120.1 F | 0.2 |
| 8:00 | 163.6 F | 0.3 |
| 10:00 | 187.4 F | 2.4 |
| 12:00 | 206.1 F | 6.0 |
| 14:00 | 217.1 F | 9.6 |
| 16:00 | 228.8 F | 13.5 |
| 18:00 | 230.7 F | 8.4 |
| 20:00 | 242.8 F | 13.5 |
| 22:00 | 241.0 F | 11.0 |
| 24:00 | 237.8 F | 9.3 |
| 26:00 | 252.9 F | 16.5 |
| 28:00 | 264.9 F | 23.6 |

STERILIZING

| mm:ss | Degrees | PSI |
|-------|---------|------|
| 0:00 | 271.5 F | 28.0 |
| 1:00 | 272.1 F | 28.3 |
| 2:00 | 272.7 F | 28.8 |
| 3:00 | 271.8 F | 28.2 |
| 4:00 | 272.0 F | 28.1 |

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING

Drying Start

| mm:ss | Degrees | PSI |
|-------|---------|-----|
| 0:00 | 195.9 F | 0.0 |
| 5:00 | 226.7 F | 0.0 |
| 10:00 | 207.4 F | 0.0 |
| 15:00 | 227.2 F | 0.0 |

Dry Cycle Aborted

FILLING: 0:55
HEATING: 29:47
STERILIZING: 4:00
VENTING: 2:51
DRYING: 18:49
TOTAL CYCLE: 00:56:23

01/27/2021 08:39 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____