

-----  
Midmark M11 - v1.0.5  
Total Cycles: 403

Name: RITTER2  
Sterilizer ID: V2313228  
09/28/2021 01:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.0 F 0.0  
2:00 195.2 F 4.2  
4:00 209.6 F 10.0  
6:00 223.1 F 6.4  
8:00 236.8 F 12.5  
10:00 239.2 F 11.2  
12:00 240.0 F 10.8  
14:00 254.3 F 18.1  
16:00 266.4 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.1 F 29.0  
2:00 272.5 F 29.3  
3:00 272.9 F 29.7  
4:00 272.0 F 29.4  
  
Min 271.8 F 28.9  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.5 F 0.0  
5:00 224.7 F 0.0  
10:00 229.7 F 0.0  
15:00 229.0 F 0.0  
20:00 228.8 F 0.0  
25:00 229.3 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 2:29  
HEATING: 17:26  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:04  
CYCLE COMPLETE

09/28/2021 02:52 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01