

-----  
Midmark M11 - v1.0.5  
Total Cycles: 925

Name: RITTER2  
Sterilizer ID: V2313228  
03/30/2022 06:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.2 F 0.0  
2:00 76.4 F 0.0  
4:00 85.3 F 0.0  
6:00 110.7 F 0.1  
8:00 159.8 F 0.6  
10:00 195.9 F 0.6  
12:00 212.1 F 2.6  
14:00 220.3 F 5.9  
16:00 230.0 F 9.4  
18:00 239.1 F 13.6  
20:00 238.5 F 10.8  
22:00 235.8 F 8.8  
24:00 240.4 F 10.0  
26:00 249.3 F 14.9  
28:00 261.0 F 21.6  
30:00 271.9 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.1  
2:00 271.3 F 28.6  
3:00 272.4 F 29.1  
4:00 271.3 F 28.4  
  
Min 271.2 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 220.2 F 0.0  
10:00 224.3 F 0.0  
15:00 231.5 F 0.0  
20:00 227.1 F 0.0  
25:00 231.2 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:33  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:29  
CYCLE COMPLETE

03/30/2022 07:50 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_