
Midmark M11 - v1.0.5
Total Cycles: 1621

Name: RITTER2
Sterilizer ID: V2313228
10/19/2022 08:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.0 F 0.0
2:00 179.8 F 2.1
4:00 179.5 F 5.8
6:00 199.8 F 10.4
8:00 215.8 F 4.7
10:00 227.7 F 9.3
12:00 236.5 F 10.8
14:00 240.5 F 11.8
16:00 239.2 F 10.5
18:00 251.9 F 16.8
20:00 262.8 F 23.2
22:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.7
2:00 271.9 F 29.1
3:00 271.7 F 29.4
4:00 272.9 F 29.7

Min 271.4 F 28.9
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 205.8 F 0.0
10:00 215.0 F 0.0
15:00 216.3 F 0.0
20:00 216.4 F 0.0
25:00 215.5 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

10/19/2022 09:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____