
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006644
9:48 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:06
STERILIZING 3:58
136.0 C 220kPa 3:58
136.2 C 223kPa 4:28
136.1 C 220kPa 4:58
136.4 C 225kPa 5:28
136.2 C 223kPa 5:58
136.3 C 222kPa 6:28
135.9 C 220kPa 6:58
136.2 C 223kPa 7:28
136.2 C 223kPa 7:28
VENTING 7:28
AIR DRYING 8:22
CYCLE COMPLETE 8:44

Digital Signature #
F2CE15A5EC223CD5
