
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008112
11:00 28/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:30
STERILIZING 3:17
136.0 C 220kPa 3:17
136.1 C 221kPa 3:47
135.9 C 221kPa 4:17
136.0 C 221kPa 4:47
136.3 C 224kPa 5:17
136.4 C 224kPa 5:47
136.4 C 222kPa 6:17
136.4 C 224kPa 6:47
136.4 C 224kPa 6:47
VENTING 6:47
AIR DRYING 7:37
CYCLE COMPLETE 37:37

Digital Signature #
576CE651C051F22E
