

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008458  
8:18 26/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:30  
STERILIZING 3:41  
136.5 C 220kPa 3:41  
136.4 C 224kPa 4:11  
136.3 C 222kPa 4:41  
136.2 C 218kPa 5:11  
135.9 C 220kPa 5:41  
136.4 C 223kPa 6:11  
136.4 C 220kPa 6:41  
136.1 C 220kPa 7:11  
136.1 C 220kPa 7:11  
VENTING 7:11  
AIR DRYING 8:01  
CYCLE COMPLETE 38:01

-----  
Digital Signature #  
D4482E226D7CC377  
-----