

-----  
Midmark M11 - v1.0.5  
Total Cycles: 552

Name: RITTER2  
Sterilizer ID: V2313228  
11/15/2021 01:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.1 F 0.0  
2:00 184.4 F 4.3  
4:00 203.6 F 10.1  
6:00 218.0 F 5.9  
8:00 235.1 F 12.3  
10:00 237.1 F 10.4  
12:00 236.9 F 9.5  
14:00 251.7 F 16.7  
16:00 264.1 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.4 F 29.4  
2:00 272.1 F 29.6  
3:00 272.1 F 29.2  
4:00 272.1 F 29.7  
  
Min 271.5 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.8 F 0.0  
5:00 230.3 F 0.0  
10:00 227.9 F 0.0  
15:00 223.1 F 0.0  
20:00 207.2 F 0.0  
25:00 202.9 F 0.0  
30:00 201.0 F 0.0  
  
Drying Complete

FILLING: 3:01  
HEATING: 17:54  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:06  
CYCLE COMPLETE

11/15/2021 02:34 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02