
Midmark M11 - v1.0.0
Total Cycles: 5114

Name: RITTER 1
Sterilizer ID: V2178198
11/20/2024 12:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.4 F 0.0
2:00 170.4 F 0.1
4:00 180.5 F 0.7
6:00 201.5 F 3.5
8:00 215.5 F 7.3
10:00 227.3 F 11.2
12:00 227.2 F 6.3
14:00 239.8 F 11.7
16:00 237.4 F 9.5
18:00 238.9 F 9.5
20:00 249.3 F 14.8
22:00 260.8 F 21.2
24:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.4 F 28.9
2:00 272.7 F 28.9
3:00 272.2 F 28.5
4:00 271.7 F 28.2

Min 271.5 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 202.0 F 0.0
10:00 201.2 F 0.0
15:00 194.9 F 0.0
20:00 188.8 F 0.0
25:00 189.5 F 0.0
30:00 191.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:47
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:03:35
CYCLE COMPLETE

11/20/2024 01:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____