
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000252
7:45 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:48
STERILIZING 5:26
136.0 C 227kPa 5:26
136.4 C 233kPa 5:56
136.4 C 233kPa 6:26
136.3 C 232kPa 6:56
136.2 C 232kPa 7:26
136.2 C 232kPa 7:56
136.4 C 231kPa 8:26
136.4 C 232kPa 8:56
136.4 C 232kPa 8:56
VENTING 8:56
AIR DRYING 9:56
CYCLE COMPLETE 21:29

Digital Signature #
A30D039B7973BD5C
