
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007619
8:12 02/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:03
PRESSURIZING 4:04
STERILIZING 5:17
136.4 C 221kPa 5:17
136.3 C 223kPa 5:47
136.3 C 223kPa 6:17
136.1 C 221kPa 6:47
136.2 C 221kPa 7:17
135.9 C 220kPa 7:47
136.4 C 221kPa 8:17
136.1 C 220kPa 8:47
136.1 C 220kPa 8:47
VENTING 8:47
AIR DRYING 9:38
CYCLE COMPLETE 37:12

Digital Signature #
3083784AA2019ADF
