
Midmark M11 - v1.0.5
Total Cycles: 3780

Name: RITTER2
Sterilizer ID: V2313228
07/15/2024 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.4 F 0.0
2:00 163.8 F 1.9
4:00 173.5 F 5.2
6:00 194.4 F 9.5
8:00 211.9 F 8.4
10:00 221.1 F 8.1
12:00 234.4 F 13.0
14:00 238.9 F 10.4
16:00 235.2 F 8.5
18:00 249.1 F 15.1
20:00 260.4 F 21.5
22:00 270.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.9 F 29.9
2:00 271.5 F 29.3
3:00 273.0 F 30.0
4:00 271.7 F 29.4

Min 271.3 F 29.2
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 231.0 F 0.0
10:00 189.5 F 0.0
15:00 181.4 F 0.0
20:00 176.6 F 0.0
25:00 174.6 F 0.0
30:00 173.3 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 23:32
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:02:27
CYCLE COMPLETE

07/15/2024 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____